

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101024\
 Data File : VV037796.D
 Acq On : 10 Oct 2024 11:50
 Operator : SY/MD
 Sample : P4363-01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOAY5

Quant Time: Oct 11 05:12:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 11 05:10:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	609932	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	601569	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	271211	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	200211	37.393	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.780%
7) Chloroethane-d5	1.532	69	171206	41.195	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.380%
11) 1,1-Dichloroethene-d2	2.059	65	91661	40.976	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.960%
21) 2-Butanone-d5	3.799	46	164328	86.168	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.170%
24) Chloroform-d	4.243	84	412842	43.339	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.680%
26) 1,2-Dichloroethane-d4	4.937	65	253296	45.428	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.860%
32) Benzene-d6	4.950	84	823621	43.753	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.500%
36) 1,2-Dichloropropane-d6	5.982	67	264150	44.068	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.140%
41) Toluene-d8	7.236	98	713926	41.825	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.660%
43) trans-1,3-Dichloroprop...	7.551	79	112167	40.282	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.560%
47) 2-Hexanone-d5	8.024	63	98834	78.084	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.080%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	294745	42.319	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.640%
66) 1,2-Dichlorobenzene-d4	11.554	152	279762	48.953	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.900%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101024\
Data File : VV037796.D
Acq On : 10 Oct 2024 11:50
Operator : SY/MD
Sample : P4363-01
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOAY5

Quant Time: Oct 11 05:12:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 11 05:10:27 2024
Response via : Initial Calibration

