

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
 Data File : VV037818.D
 Acq On : 11 Oct 2024 10:30
 Operator : SY/MD
 Sample : P4363-03
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 14 06:16:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 14 06:04:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	648847	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	648625	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	284887	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	193417	33.958	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.920%
7) Chloroethane-d5	1.532	69	170783	38.628	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.260%
11) 1,1-Dichloroethene-d2	2.060	65	90623	38.082	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.160%
21) 2-Butanone-d5	3.793	46	150586	74.226	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	74.230%
24) Chloroform-d	4.239	84	416397	41.091	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.180%
26) 1,2-Dichloroethane-d4	4.934	65	258709	43.616	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.240%
32) Benzene-d6	4.950	84	823685	40.582	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.160%
36) 1,2-Dichloropropane-d6	5.982	67	267968	41.461	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.920%
41) Toluene-d8	7.236	98	722845	39.275	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.560%#
43) trans-1,3-Dichloroprop...	7.548	79	109970	36.628	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.260%
47) 2-Hexanone-d5	8.024	63	90834	66.558	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	66.560%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	290099	38.630	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.260%
66) 1,2-Dichlorobenzene-d4	11.551	152	280002	46.643	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
Data File : VV037818.D
Acq On : 11 Oct 2024 10:30
Operator : SY/MD
Sample : P4363-03
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 14 06:16:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 14 06:04:26 2024
Response via : Initial Calibration

