

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
 Data File : VV037824.D
 Acq On : 11 Oct 2024 12:39
 Operator : SY/MD
 Sample : P4346-03ME
 Misc : 3.45G/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOAW5ME

Quant Time: Oct 14 06:17:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 14 06:04:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	664852	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	666487	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	339647	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	200374	34.332	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.660%
7) Chloroethane-d5	1.503	69	80125	17.687	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	35.380%#
11) 1,1-Dichloroethene-d2	2.043	65	97613	40.032	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.060%
21) 2-Butanone-d5	3.809	46	192997	92.841	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.840%
24) Chloroform-d	4.236	84	451536	43.486	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.980%
26) 1,2-Dichloroethane-d4	4.934	65	283177	46.592	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.180%
32) Benzene-d6	4.947	84	899505	43.130	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.260%
36) 1,2-Dichloropropane-d6	5.982	67	290409	43.729	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.460%
41) Toluene-d8	7.236	98	790554	41.803	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	7.548	79	125490	40.677	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.360%
47) 2-Hexanone-d5	8.030	63	131872	94.038	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.040%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	347017	44.971	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.940%
66) 1,2-Dichlorobenzene-d4	11.554	152	340261	47.542	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
Data File : VV037824.D
Acq On : 11 Oct 2024 12:39
Operator : SY/MD
Sample : P4346-03ME
Misc : 3.45G/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOAW5ME

Quant Time: Oct 14 06:17:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 14 06:04:26 2024
Response via : Initial Calibration

