

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
 Data File : VV037826.D
 Acq On : 11 Oct 2024 13:22
 Operator : SY/MD
 Sample : P4346-11ME
 Misc : 3.28G/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOAX1ME

Quant Time: Oct 14 06:17:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 14 06:04:26 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/14/2024
 Supervised By :Mahesh Dadoda 10/14/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	577325	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	572082	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	276106	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	230388	45.459	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	90.920%		
7) Chloroethane-d5	1.503	69	93925	23.876	ug/L	-0.03
Spiked Amount 50.000	Range 70 - 130		Recovery =	47.760%#		
11) 1,1-Dichloroethene-d2	2.044	65	108743	51.358	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery =	102.720%		
21) 2-Butanone-d5	3.806	46	191575	106.129	ug/L	0.02
Spiked Amount 100.000	Range 40 - 130		Recovery =	106.130%		
24) Chloroform-d	4.240	84	508256m	56.369	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	112.740%		
26) 1,2-Dichloroethane-d4	4.931	65	318580	60.364	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	120.720%		
32) Benzene-d6	4.947	84	1015203	56.710	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	113.420%		
36) 1,2-Dichloropropane-d6	5.982	67	329560	57.814	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	115.620%		
41) Toluene-d8	7.236	98	900695	55.487	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	110.980%		
43) trans-1,3-Dichloroprop...	7.548	79	141918	53.593	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	107.180%		
47) 2-Hexanone-d5	8.030	63	166093	137.987	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	137.990%#		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	397678	60.041	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	120.080%#		
66) 1,2-Dichlorobenzene-d4	11.554	152	375474	64.535	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	129.080%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
Data File : VV037826.D
Acq On : 11 Oct 2024 13:22
Operator : SY/MD
Sample : P4346-11ME
Misc : 3.28G/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOAX1ME

Quant Time: Oct 14 06:17:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 14 06:04:26 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/14/2024
Supervised By : Mahesh Dadoda 10/14/2024

