

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
 Data File : VV037816.D
 Acq On : 11 Oct 2024 09:39
 Operator : SY/MD
 Sample : VV1011MBL01
 Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK240

Quant Time: Oct 14 06:15:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 14 06:04:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	588695	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	586887	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	270708	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	206780	40.013	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.020%
7) Chloroethane-d5	1.536	69	183018	45.626	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.260%
11) 1,1-Dichloroethene-d2	2.060	65	95210	44.098	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.200%
21) 2-Butanone-d5	3.793	46	167196	90.834	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.830%
24) Chloroform-d	4.240	84	448846	48.819	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.640%
26) 1,2-Dichloroethane-d4	4.934	65	267284	49.667	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.340%
32) Benzene-d6	4.950	84	853996	46.502	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	5.982	67	282814	48.362	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.720%
41) Toluene-d8	7.236	98	738865	44.369	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.740%
43) trans-1,3-Dichloroprop...	7.548	79	115601	42.554	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.100%
47) 2-Hexanone-d5	8.024	63	96061	77.792	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.790%
56) 1,1,2,2-Tetrachloroeth...	10.143	84	305087	44.900	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	11.551	152	292598	51.294	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
Data File : VV037816.D
Acq On : 11 Oct 2024 09:39
Operator : SY/MD
Sample : VV1011MBL01
Misc : 5.00g/5mL/100ul/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK240

Quant Time: Oct 14 06:15:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 14 06:04:26 2024
Response via : Initial Calibration

