

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
 Data File : VV037817.D
 Acq On : 11 Oct 2024 10:00
 Operator : SY/MD
 Sample : VV1011WBL01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK241

Quant Time: Oct 14 06:15:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 14 06:04:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	561437	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	563124	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	257263	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	196609	39.892	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.780%
7) Chloroethane-d5	1.532	69	176376	46.104	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.060	65	92212	44.783	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.560%
21) 2-Butanone-d5	3.796	46	160629	91.503	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.500%
24) Chloroform-d	4.243	84	426249	48.612	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.220%
26) 1,2-Dichloroethane-d4	4.937	65	258127	50.294	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.580%
32) Benzene-d6	4.953	84	825804	46.864	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.720%
36) 1,2-Dichloropropane-d6	5.982	67	272556	48.574	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.140%
41) Toluene-d8	7.236	98	702674	43.976	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.960%
43) trans-1,3-Dichloroprop...	7.548	79	107907	41.398	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.800%
47) 2-Hexanone-d5	8.024	63	94688	79.916	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.920%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	304033	46.633	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.260%
66) 1,2-Dichlorobenzene-d4	11.551	152	281813	51.985	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
Data File : VV037817.D
Acq On : 11 Oct 2024 10:00
Operator : SY/MD
Sample : VV1011WBL01
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK241

Quant Time: Oct 14 06:15:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 14 06:04:26 2024
Response via : Initial Calibration

