

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
 Data File : VV037823.D
 Acq On : 11 Oct 2024 12:17
 Operator : SY/MD
 Sample : P4363-03
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 14 06:17:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 14 06:04:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	587824	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	589165	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	271460	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	186066	36.058	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.120%
7) Chloroethane-d5	1.532	69	167041	41.704	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.400%
11) 1,1-Dichloroethene-d2	2.060	65	89540	41.533	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.060%
21) 2-Butanone-d5	3.799	46	170075	92.535	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.540%
24) Chloroform-d	4.243	84	417835	45.513	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.020%
26) 1,2-Dichloroethane-d4	4.937	65	254036	47.275	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.540%
32) Benzene-d6	4.953	84	794766	43.109	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.220%
36) 1,2-Dichloropropane-d6	5.985	67	264908	45.124	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.240%
41) Toluene-d8	7.239	98	674817	40.366	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.740%
43) trans-1,3-Dichloroprop...	7.551	79	109427	40.125	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.260%
47) 2-Hexanone-d5	8.024	63	104748	84.499	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.500%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	315447	46.245	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.480%
66) 1,2-Dichlorobenzene-d4	11.554	152	276598	48.355	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
Data File : VV037823.D
Acq On : 11 Oct 2024 12:17
Operator : SY/MD
Sample : P4363-03
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 14 06:17:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 14 06:04:26 2024
Response via : Initial Calibration

