

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
 Data File : VV037831.D
 Acq On : 11 Oct 2024 15:10
 Operator : SY/MD
 Sample : VIBLK203
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK203

Quant Time: Oct 14 06:19:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 14 06:04:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	692222	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	659907	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	312726	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	192409	31.664	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.320%
7) Chloroethane-d5	1.532	69	166085	35.212	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.420%
11) 1,1-Dichloroethene-d2	2.060	65	89694	35.330	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.660%
21) 2-Butanone-d5	3.799	46	174624	80.681	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.680%
24) Chloroform-d	4.243	84	421274	38.967	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.940%
26) 1,2-Dichloroethane-d4	4.937	65	263011	41.563	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.120%
32) Benzene-d6	4.950	84	844421	40.892	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.780%
36) 1,2-Dichloropropane-d6	5.985	67	269140	40.931	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	81.860%
41) Toluene-d8	7.236	98	734166	39.209	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.420%#
43) trans-1,3-Dichloroprop...	7.548	79	112606	36.865	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.720%
47) 2-Hexanone-d5	8.024	63	109104	78.578	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.580%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	304373	39.838	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	79.680%
66) 1,2-Dichlorobenzene-d4	11.554	152	290684	44.111	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.220%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101124\
Data File : VV037831.D
Acq On : 11 Oct 2024 15:10
Operator : SY/MD
Sample : VIBLK203
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK203

Quant Time: Oct 14 06:19:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 14 06:04:26 2024
Response via : Initial Calibration

