

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101524\
 Data File : VV037864.D
 Acq On : 15 Oct 2024 10:59
 Operator : SY/MD
 Sample : VV1015WBL02
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK246

Quant Time: Oct 16 03:30:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 16 03:26:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	393141	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	396854	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	181963	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	273141	79.145	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 158.280%#		
7) Chloroethane-d5	1.532	69	230160	85.918	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 171.840%#		
11) 1,1-Dichloroethene-d2	2.059	65	122731	85.120	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 170.240%#		
21) 2-Butanone-d5	3.792	46	167566	136.317	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 136.320%#		
24) Chloroform-d	4.236	84	514772	83.839	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 167.680%#		
26) 1,2-Dichloroethane-d4	4.931	65	306157	85.188	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 170.380%#		
32) Benzene-d6	4.947	84	1043116	83.998	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 168.000%#		
36) 1,2-Dichloropropane-d6	5.979	67	327016	82.697	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 165.400%#		
41) Toluene-d8	7.233	98	906758	80.525	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 161.040%#		
43) trans-1,3-Dichloroprop...	7.545	79	133781	72.827	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 145.660%#		
47) 2-Hexanone-d5	8.021	63	97615	116.904	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 116.900%		
56) 1,1,2,2-Tetrachloroeth...	10.143	84	324095	70.537	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 141.080%#		
66) 1,2-Dichlorobenzene-d4	11.551	152	330922	86.305	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 172.620%#		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101524\
Data File : VV037864.D
Acq On : 15 Oct 2024 10:59
Operator : SY/MD
Sample : VV1015WBL02
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK246

Quant Time: Oct 16 03:30:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 16 03:26:57 2024
Response via : Initial Calibration

