

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101524\
 Data File : VV037862.D
 Acq On : 15 Oct 2024 09:59
 Operator : SY/MD
 Sample : VV1015MBL01
 Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK244

Quant Time: Oct 16 03:30:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 16 03:26:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	633083	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	628210	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	285684	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	232863	41.901	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.800%
7) Chloroethane-d5	1.535	69	194014	44.976	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.960%
11) 1,1-Dichloroethene-d2	2.060	65	103503	44.578	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.160%
21) 2-Butanone-d5	3.793	46	142289	71.883	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	71.880%
24) Chloroform-d	4.240	84	435971	44.094	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.180%
26) 1,2-Dichloroethane-d4	4.934	65	260743	45.054	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.100%
32) Benzene-d6	4.950	84	876325	44.579	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.160%
36) 1,2-Dichloropropane-d6	5.979	67	278390	44.474	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.940%
41) Toluene-d8	7.233	98	756468	42.438	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.880%
43) trans-1,3-Dichloroprop...	7.548	79	111573	38.369	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.740%
47) 2-Hexanone-d5	8.024	63	83158	62.913	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	62.910%
56) 1,1,2,2-Tetrachloroeth...	10.143	84	275190	37.836	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	75.680%
66) 1,2-Dichlorobenzene-d4	11.551	152	287153	47.700	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101524\
Data File : VV037862.D
Acq On : 15 Oct 2024 09:59
Operator : SY/MD
Sample : VV1015MBL01
Misc : 5.00g/5mL/100ul/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK244

Quant Time: Oct 16 03:30:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 16 03:26:57 2024
Response via : Initial Calibration

