

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101624\
 Data File : VV037875.D
 Acq On : 16 Oct 2024 13:38
 Operator : SY/MD
 Sample : VIBLK388
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK388

Quant Time: Oct 17 02:47:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 17 02:37:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	477107	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	484944	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	225454	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	89602	2.851	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.000%
7) Chloroethane-d5	1.529	69	80915	2.891	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	57.800%#
11) 1,1-Dichloroethene-d2	2.053	65	41974	2.868	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.400%#
20) 2-Butanone-d5	3.825	46	190030	26.400	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	52.800%
24) Chloroform-d	4.236	84	185393	2.750	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	55.000%#
26) 1,2-Dichloroethane-d4	4.934	65	87900	2.825	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	56.600%#
32) Benzene-d6	4.950	84	363782	2.934	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	58.600%#
36) 1,2-Dichloropropane-d6	5.979	67	112801	3.043	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	60.800%
41) Toluene-d8	7.236	98	313576	2.755	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	55.000%#
43) trans-1,3-Dichloroprop...	7.558	79	37483	2.642	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	52.800%#
46) 2-Hexanone-d5	8.024	63	149037	25.100	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	50.200%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	82465	2.716	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	54.400%#
66) 1,2-Dichlorobenzene-d4	11.554	152	117441	3.055	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	61.000%#
Target Compounds						Qvalue
70) 1,2,4-trichlorobenzene	13.201	180	10944	0.255	ug/L	# 86
72) 1,2,3-Trichlorobenzene	13.680	180	13222	0.348	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101624\
Data File : VV037875.D
Acq On : 16 Oct 2024 13:38
Operator : SY/MD
Sample : VIBLK388
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK388

Quant Time: Oct 17 02:47:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 17 02:37:45 2024
Response via : Initial Calibration

