

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101722\
 Data File : VV028584.D
 Acq On : 17 Oct 2022 13:21
 Operator : SY/MD
 Sample : N5026-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 18 01:22:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 18 01:19:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	187386	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	177686	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	84075	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.297	65	85198	5.283	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.600%
7) Chloroethane-d5	1.561	69	79397	5.989	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.800%
11) 1,1-Dichloroethene-d2	2.098	63	116914	4.003	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.000%
20) 2-Butanone-d5	3.963	46	190026	54.811	ug/L	0.08
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.620%
24) Chloroform-d	4.346	84	137957	5.587	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.800%
26) 1,2-Dichloroethane-d4	5.034	65	76169	5.918	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.400%
32) Benzene-d6	5.046	84	282869	4.930	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
36) 1,2-Dichloropropane-d6	6.069	67	98489	5.319	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.400%
41) Toluene-d8	7.313	98	234592	5.161	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
43) trans-1,3-Dichloroprop...	7.622	79	29621	4.736	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.800%
46) 2-Hexanone-d5	8.091	63	170203	63.047	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	126.100%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	65822	5.970	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	119.400%
66) 1,2-Dichlorobenzene-d4	11.619	152	80242	5.784	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101722\
Data File : VV028584.D
Acq On : 17 Oct 2022 13:21
Operator : SY/MD
Sample : N5026-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 18 01:22:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 18 01:19:01 2022
Response via : Initial Calibration

