

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101722\
 Data File : VV028599.D
 Acq On : 17 Oct 2022 19:44
 Operator : SY/MD
 Sample : N5103-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BA1

Quant Time: Oct 18 02:51:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 18 01:19:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	172337	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	169610	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	79940	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	68151	4.595	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.800%
7) Chloroethane-d5	1.564	69	65512	5.373	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.400%
11) 1,1-Dichloroethene-d2	2.105	63	97558	3.632	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.600%
20) 2-Butanone-d5	3.908	46	111602	35.001	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	70.000%
24) Chloroform-d	4.346	84	116630	5.135	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
26) 1,2-Dichloroethane-d4	5.030	65	59897	5.060	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.047	84	234515	4.282	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	6.066	67	81452	4.609	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.200%
41) Toluene-d8	7.313	98	195565	4.507	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
43) trans-1,3-Dichloroprop...	7.622	79	23750	3.978	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.088	63	133478	51.797	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.600%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	55975	5.318	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.400%
66) 1,2-Dichlorobenzene-d4	11.615	152	66071	5.009	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101722\
Data File : VV028599.D
Acq On : 17 Oct 2022 19:44
Operator : SY/MD
Sample : N5103-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BA1

Quant Time: Oct 18 02:51:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 18 01:19:01 2022
Response via : Initial Calibration

