

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101722\
 Data File : VW028587.D
 Acq On : 17 Oct 2022 14:33
 Operator : SY/MD
 Sample : N5144-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BT8

Quant Time: Oct 18 01:23:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 18 01:19:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	189192	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	179622	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	84451	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	74943	4.603	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.564	69	73333	5.478	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.600%
11) 1,1-Dichloroethene-d2	2.105	63	112591	3.818	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.400%
20) 2-Butanone-d5	3.912	46	177055	50.582	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.160%
24) Chloroform-d	4.346	84	134313	5.387	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.800%
26) 1,2-Dichloroethane-d4	5.031	65	68578	5.277	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.047	84	268081	4.622	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	6.066	67	95413	5.098	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.000%
41) Toluene-d8	7.313	98	225577	4.909	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
43) trans-1,3-Dichloroprop...	7.622	79	29133	4.607	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.200%
46) 2-Hexanone-d5	8.088	63	155704	57.054	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.100%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	62660	5.622	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.400%
66) 1,2-Dichlorobenzene-d4	11.615	152	75775	5.438	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101722\
Data File : VV028587.D
Acq On : 17 Oct 2022 14:33
Operator : SY/MD
Sample : N5144-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BT8

Quant Time: Oct 18 01:23:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 18 01:19:01 2022
Response via : Initial Calibration

