

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101722\
 Data File : VV028596.D
 Acq On : 17 Oct 2022 18:32
 Operator : SY/MD
 Sample : N5103-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BA2

Quant Time: Oct 18 02:51:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 18 01:19:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	182950	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	174312	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	80254	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.298	65	80818	5.133	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.600%
7) Chloroethane-d5	1.561	69	72700	5.616	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.400%
11) 1,1-Dichloroethene-d2	2.101	63	108971	3.821	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.400%
20) 2-Butanone-d5	3.934	46	172985	51.106	ug/L	0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.220%
24) Chloroform-d	4.342	84	124757	5.175	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.400%
26) 1,2-Dichloroethane-d4	5.031	65	68329	5.438	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.800%
32) Benzene-d6	5.047	84	252390	4.484	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
36) 1,2-Dichloropropane-d6	6.066	67	88781	4.888	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.310	98	208490	4.675	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
43) trans-1,3-Dichloroprop...	7.619	79	26718	4.354	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.088	63	145207	54.829	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.660%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	58960	5.451	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.000%
66) 1,2-Dichlorobenzene-d4	11.619	152	70304	5.309	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101722\
Data File : VV028596.D
Acq On : 17 Oct 2022 18:32
Operator : SY/MD
Sample : N5103-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BA2

Quant Time: Oct 18 02:51:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 18 01:19:01 2022
Response via : Initial Calibration

