

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102324\
 Data File : VV037914.D
 Acq On : 23 Oct 2024 11:21
 Operator : SY/MD
 Sample : VV1023MBL01
 Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK253

Quant Time: Oct 24 01:50:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 22 05:07:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	456976	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	455479	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	210800	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	167651	41.792	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	83.580%	
7) Chloroethane-d5	1.532	69	150753	48.415	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.820%	
11) 1,1-Dichloroethene-d2	2.060	65	79272	47.299	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	94.600%	
21) 2-Butanone-d5	3.799	46	166866	116.785	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	116.790%	
24) Chloroform-d	4.243	84	385138	53.964	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	107.920%	
26) 1,2-Dichloroethane-d4	4.937	65	239398	57.307	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	114.620%	
32) Benzene-d6	4.953	84	744173	52.212	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.420%	
36) 1,2-Dichloropropane-d6	5.982	67	238849	52.627	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	105.260%	
41) Toluene-d8	7.236	98	635046	49.137	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	98.280%	
43) trans-1,3-Dichloroprop...	7.551	79	101613	48.196	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	96.400%	
47) 2-Hexanone-d5	8.024	63	102215	106.657	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	106.660%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	288507	54.710	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	109.420%	
66) 1,2-Dichlorobenzene-d4	11.554	152	250990	56.504	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	113.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV102324\
Data File : VV037914.D
Acq On : 23 Oct 2024 11:21
Operator : SY/MD
Sample : VV1023MBL01
Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK253

Quant Time: Oct 24 01:50:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 22 05:07:16 2024
Response via : Initial Calibration

