

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102523\
 Data File : VV032622.D
 Acq On : 25 Oct 2023 10:47
 Operator : SY/MD
 Sample : VV1025WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK380

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/26/2023
 Supervised By : Mahesh Dadoda 10/27/2023

Quant Time: Oct 26 05:00:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 26 04:59:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	85535	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	96844	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	41051	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	35028	5.035	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 100.800%		
7) Chloroethane-d5	1.532	69	30950	5.481	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 109.600%		
11) 1,1-Dichloroethene-d2	2.060	65	14273	4.865	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 97.400%		
20) 2-Butanone-d5	3.818	46	58791m	33.409	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 66.820%		
24) Chloroform-d	4.253	84	57265	4.666	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 93.400%		
26) 1,2-Dichloroethane-d4	4.947	65	28600	4.920	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 98.400%		
32) Benzene-d6	4.960	84	116416	4.085	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 81.800%		
36) 1,2-Dichloropropane-d6	5.992	67	36595	3.876	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 77.600%		
41) Toluene-d8	7.246	98	107213	4.114	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 82.200%		
43) trans-1,3-Dichloroprop...	7.561	79	13504	4.036	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 80.800%		
46) 2-Hexanone-d5	8.031	63	54563	34.662	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 69.320%		
56) 1,1,2,2-Tetrachloroeth...	10.156	84	28522	4.319	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 86.400%		
66) 1,2-Dichlorobenzene-d4	11.564	152	38005	4.989	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 99.800%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102523\
Data File : VV032622.D
Acq On : 25 Oct 2023 10:47
Operator : SY/MD
Sample : VV1025WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK380

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/26/2023
Supervised By : Mahesh Dadoda 10/27/2023

Quant Time: Oct 26 05:00:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 26 04:59:52 2023
Response via : Initial Calibration

