



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102523\
Data File : VV032620.D
Acq On : 25 Oct 2023 09:25
Operator : SY/MD
Sample : BFB227
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB227

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Thu Oct 26 04:59:52 2023



AutoFind: Scans 2790, 2791, 2792; Background Corrected with Scan 2778

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	16405	PASS
75	95	30	80	54.7	43560	PASS
95	95	100	100	100.0	79664	PASS
96	95	5	9	7.2	5707	PASS
173	174	0.00	2	1.3	846	PASS
174	95	50	120	84.0	66931	PASS
175	174	5	9	7.7	5153	PASS
176	174	95	105	96.1	64317	PASS
177	176	5	10	6.4	4105	PASS

m/z	Abundance
35.90	113.0
38.30	104.0
39.80	156.0
43.90	306.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

m/z	Abundance
35.90	914.0
37.00	4646.0
38.00	4356.0
39.00	2067.0
39.80	288.0
42.80	163.0
43.90	856.0
44.90	752.0
46.00	179.0
46.90	976.0
48.00	585.0
49.00	3860.0
50.00	16632.0
51.00	5444.0
51.90	258.0
53.00	185.0
53.90	111.0
54.90	344.0
55.90	1396.0
57.00	2711.0
57.90	156.0
60.00	937.0
61.00	4655.0
62.00	4612.0
62.90	3868.0
64.00	605.0
65.00	801.0
65.90	138.0
67.00	284.0
68.00	8740.0
69.00	9190.0
69.90	779.0
71.90	541.0
73.00	4052.0
74.00	15579.0
75.00	44536.0
76.00	3653.0
76.90	667.0
77.80	413.0
78.90	3929.0
79.90	1042.0
80.90	3798.0
81.90	839.0
82.90	141.0
86.90	3595.0
87.90	3583.0
91.00	400.0
91.90	2556.0
93.00	4054.0
94.00	11038.0
95.00	79744.0
96.00	6006.0
97.00	228.0
102.80	106.0
103.80	439.0
104.70	152.0
105.90	430.0
107.00	140.0
110.70	151.0
112.00	101.0
112.90	158.0
114.80	143.0
115.90	330.0
116.90	647.0
117.80	367.0
118.80	596.0
127.90	385.0
128.80	178.0
129.70	337.0
130.90	126.0
134.80	185.0
136.80	201.0
139.90	117.0
140.80	984.0
141.80	197.0
142.80	1414.0
143.80	144.0
144.90	359.0
145.70	155.0
147.80	176.0
150.00	137.0
154.90	250.0
156.90	171.0
158.80	116.0
160.90	135.0
171.90	1100.0
173.00	1185.0
173.90	66056.0
175.00	5246.0
175.90	62736.0
176.90	4168.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

m/z	Abundance
35.90	933.0
37.00	4689.0
38.00	4412.0
39.00	2034.0
39.80	290.0
42.20	108.0
42.80	133.0
43.90	928.0
44.90	747.0
46.00	201.0
46.90	999.0
48.00	561.0
49.00	3933.0
50.00	16936.0
51.00	5486.0
51.90	263.0
53.00	171.0
53.90	102.0
54.90	365.0
55.90	1387.0
57.00	2841.0
57.90	160.0
59.10	109.0
60.00	899.0
61.00	4494.0
62.00	4732.0
63.00	3963.0
63.90	628.0
64.90	854.0
66.00	138.0
67.00	300.0
68.00	8949.0
69.00	9309.0
69.90	733.0
71.90	548.0
73.00	4079.0
74.00	15860.0
75.00	44984.0
76.00	3729.0
76.90	685.0
77.90	424.0
78.90	3949.0
79.90	1126.0
80.90	3711.0
81.90	793.0
82.90	134.0
86.90	3775.0
87.90	3645.0
90.90	453.0
91.90	2729.0
93.00	4153.0
94.00	11458.0
95.00	82328.0
96.00	5933.0
97.00	206.0
102.80	106.0
103.80	412.0
104.90	153.0
105.90	429.0
107.00	131.0
110.70	112.0
110.90	112.0
112.00	102.0
112.90	144.0
114.80	173.0
115.90	370.0
116.90	665.0
117.80	366.0
118.80	546.0
127.80	356.0
128.80	191.0
129.80	355.0
130.80	118.0
134.80	209.0
136.80	179.0
139.90	127.0
140.80	987.0
141.80	210.0
142.80	1409.0
143.80	118.0
144.90	351.0
145.70	184.0
147.80	187.0
149.90	131.0
154.90	302.0
157.00	149.0
158.80	135.0
160.90	147.0
170.60	102.0
171.90	1091.0
173.00	1354.0
173.90	69320.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

175.00	5322.0
175.90	66568.0
176.90	4230.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

m/z	Abundance
35.90	893.0
37.00	4298.0
38.00	4157.0
39.00	1803.0
39.90	280.0
40.60	118.0
42.10	108.0
42.80	113.0
43.90	915.0
44.90	674.0
46.00	205.0
46.90	849.0
48.00	556.0
49.00	3679.0
50.00	15648.0
51.00	5073.0
51.90	243.0
52.90	152.0
55.00	364.0
55.90	1314.0
57.00	2529.0
58.00	132.0
59.90	756.0
61.00	3941.0
62.00	4446.0
63.00	3652.0
63.90	570.0
64.90	897.0
65.80	128.0
67.00	265.0
68.00	8331.0
69.00	8577.0
69.90	620.0
71.90	500.0
73.00	3777.0
74.00	14665.0
75.00	41160.0
76.00	3511.0
76.80	612.0
77.90	394.0
78.90	3526.0
79.90	1064.0
80.90	3318.0
81.90	729.0
83.00	111.0
86.10	220.0
86.90	3405.0
87.90	3360.0
90.90	468.0
91.90	2517.0
93.00	3723.0
94.00	10438.0
95.00	76920.0
96.00	5183.0
96.90	165.0
102.80	109.0
103.90	372.0
104.90	146.0
105.90	367.0
112.00	101.0
112.90	107.0
114.80	176.0
115.90	388.0
116.90	636.0
117.70	322.0
118.80	468.0
127.90	306.0
128.80	178.0
129.80	337.0
130.70	100.0
134.80	198.0
136.80	175.0
139.90	117.0
140.80	888.0
141.80	169.0
142.90	1289.0
144.90	316.0
145.70	173.0
147.80	180.0
149.90	109.0
154.90	284.0
157.00	155.0
158.80	135.0
160.90	134.0
170.50	113.0
171.90	1030.0
173.90	65416.0
175.00	4891.0
175.90	63648.0
176.90	3916.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227