

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110724\
 Data File : VV037953.D
 Acq On : 07 Nov 2024 13:27
 Operator : SY/MD
 Sample : P4729-04DL 100X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B87DL

Quant Time: Nov 08 01:28:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 08 01:25:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	258632	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	275543	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	121606	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	98489	4.975	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.600%
7) Chloroethane-d5	1.529	69	78675	4.876	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.600%
11) 1,1-Dichloroethene-d2	2.053	65	42170	4.909	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.200%
20) 2-Butanone-d5	3.809	46	179534	48.658	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.320%
24) Chloroform-d	4.243	84	187790	5.204	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	4.941	65	88570	5.492	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.800%
32) Benzene-d6	4.953	84	357069	5.025	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
36) 1,2-Dichloropropane-d6	5.982	67	109857	5.294	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.800%
41) Toluene-d8	7.240	98	293409	4.594	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.800%
43) trans-1,3-Dichloroprop...	7.558	79	36533	4.553	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%
46) 2-Hexanone-d5	8.027	63	163769	52.521	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.040%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	90222	5.393	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	112304	5.599	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.000%
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	3.812	96	4733	0.235	ug/L	95
34) Trichloroethene	5.828	95	171056	8.158	ug/L	99
47) Tetrachloroethene	7.905	164	4158	0.255	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110724\
Data File : VV037953.D
Acq On : 07 Nov 2024 13:27
Operator : SY/MD
Sample : P4729-04DL 100X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B87DL

Quant Time: Nov 08 01:28:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 08 01:25:43 2024
Response via : Initial Calibration

