

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110724\
 Data File : VV037954.D
 Acq On : 07 Nov 2024 14:44
 Operator : SY/MD
 Sample : VIBLK393
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK393

Quant Time: Nov 08 01:28:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 08 01:25:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	235676	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	252094	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	109967	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	117006	6.486	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 129.800%	
7) Chloroethane-d5	1.529	69	101274	6.888	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 137.800%#	
11) 1,1-Dichloroethene-d2	2.053	65	51487	6.577	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 131.600%#	
20) 2-Butanone-d5	3.796	46	230127	68.446	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	= 136.900%#	
24) Chloroform-d	4.240	84	224476	6.827	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 136.600%#	
26) 1,2-Dichloroethane-d4	4.934	65	104099	7.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 141.600%#	
32) Benzene-d6	4.950	84	412232	6.341	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 126.800%#	
36) 1,2-Dichloropropane-d6	5.979	67	130335	6.865	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 137.400%	
41) Toluene-d8	7.236	98	335113	5.736	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 114.800%	
43) trans-1,3-Dichloroprop...	7.555	79	39825	5.425	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	= 108.400%	
46) 2-Hexanone-d5	8.021	63	204903	71.825	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 143.660%#	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	109862	7.177	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 143.600%#	
66) 1,2-Dichlorobenzene-d4	11.554	152	129656	7.148	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 143.000%#	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.442	84	31849	1.695	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110724\
Data File : VV037954.D
Acq On : 07 Nov 2024 14:44
Operator : SY/MD
Sample : VIBLK393
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK393

Quant Time: Nov 08 01:28:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 08 01:25:43 2024
Response via : Initial Calibration

