

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110724\
 Data File : VV037966.D
 Acq On : 07 Nov 2024 19:35
 Operator : SY/MD
 Sample : P4715-15
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AL3

Quant Time: Nov 08 01:31:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 08 01:25:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	269616	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	288854	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	126559	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	92036	4.460	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.200%
7) Chloroethane-d5	1.529	69	73685	4.381	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	2.053	65	40481	4.520	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.400%
20) 2-Butanone-d5	3.802	46	166061	43.173	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.340%
24) Chloroform-d	4.243	84	178565	4.747	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
26) 1,2-Dichloroethane-d4	4.940	65	84656	5.035	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	4.953	84	327406	4.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
36) 1,2-Dichloropropane-d6	5.985	67	106184	4.881	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.239	98	280677	4.193	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.800%
43) trans-1,3-Dichloroprop...	7.561	79	30558	3.633	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.600%
46) 2-Hexanone-d5	8.024	63	157365	48.142	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.280%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	90097	5.137	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	109333	5.238	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110724\
Data File : VV037966.D
Acq On : 07 Nov 2024 19:35
Operator : SY/MD
Sample : P4715-15
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COAL3

Quant Time: Nov 08 01:31:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 08 01:25:43 2024
Response via : Initial Calibration

