

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110824\
 Data File : VV037978.D
 Acq On : 08 Nov 2024 14:59
 Operator : SY/MD
 Sample : P4749-21
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GCP94

Quant Time: Nov 09 02:23:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 09 02:20:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	253854	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	275329	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	126001	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	85898	4.421	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.529	69	70026	4.422	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.400%
11) 1,1-Dichloroethene-d2	2.053	65	38276	4.539	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.800%
20) 2-Butanone-d5	3.828	46	175394	48.431	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.860%
24) Chloroform-d	4.243	84	172150	4.860	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	4.937	65	81335	5.138	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.800%
32) Benzene-d6	4.953	84	324019	4.563	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
36) 1,2-Dichloropropane-d6	5.986	67	103306	4.982	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.600%
41) Toluene-d8	7.240	98	278919	4.371	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.558	79	31577	3.938	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.800%
46) 2-Hexanone-d5	8.024	63	164594	52.827	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.660%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	88608	5.300	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.000%
66) 1,2-Dichlorobenzene-d4	11.554	152	109774	5.282	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.600%
Target Compounds						
					Qvalue	
13) Acetone	2.147	43	43028	16.738	ug/L	96
25) Chloroform	4.265	83	13272	0.371	ug/L	93
42) Toluene	7.317	91	72248	0.902	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110824\
Data File : VV037978.D
Acq On : 08 Nov 2024 14:59
Operator : SY/MD
Sample : P4749-21
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GCP94

Quant Time: Nov 09 02:23:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 09 02:20:30 2024
Response via : Initial Calibration

