

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110824\
 Data File : VV037971.D
 Acq On : 08 Nov 2024 12:06
 Operator : SY/MD
 Sample : P4729-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 09 02:21:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 09 02:20:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	258825	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	283567	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	124624	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	95162	4.803	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.000%
7) Chloroethane-d5	1.529	69	79906	4.949	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.050	65	43230	5.028	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.600%
20) 2-Butanone-d5	3.834	46	180889	48.989	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.980%
24) Chloroform-d	4.239	84	187658	5.196	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	4.937	65	89133	5.523	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.400%
32) Benzene-d6	4.950	84	352782	4.824	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
36) 1,2-Dichloropropane-d6	5.982	67	108201	5.067	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.236	98	292956	4.458	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
43) trans-1,3-Dichloroprop...	7.558	79	35855	4.342	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.800%
46) 2-Hexanone-d5	8.027	63	165836	51.679	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.360%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	89704	5.210	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.200%
66) 1,2-Dichlorobenzene-d4	11.554	152	113668	5.530	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110824\
Data File : VV037971.D
Acq On : 08 Nov 2024 12:06
Operator : SY/MD
Sample : P4729-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 09 02:21:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 09 02:20:30 2024
Response via : Initial Calibration

