

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
 Data File : VV038026.D
 Acq On : 12 Nov 2024 18:39
 Operator : SY/MD
 Sample : P4761-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A0ALO

Quant Time: Nov 13 00:49:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	274052	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	280498	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	130393	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	79822	3.346	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.000%
7) Chloroethane-d5	1.529	69	68136	3.504	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.000%
11) 1,1-Dichloroethene-d2	2.053	65	36414	3.484	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.600%
20) 2-Butanone-d5	3.802	46	192105	35.045	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	70.100%
24) Chloroform-d	4.239	84	170525	3.793	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.800%
26) 1,2-Dichloroethane-d4	4.937	65	85261	3.993	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
32) Benzene-d6	4.950	84	322944	3.726	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.600%
36) 1,2-Dichloropropane-d6	5.982	67	100719	3.710	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	74.200%
41) Toluene-d8	7.236	98	278057	3.508	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.200%
43) trans-1,3-Dichloroprop...	7.558	79	34219	3.109	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.200%
46) 2-Hexanone-d5	8.021	63	171299	34.817	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.640%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	86059	3.832	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.600%
66) 1,2-Dichlorobenzene-d4	11.554	152	113606	4.190	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.800%
Target Compounds					Qvalue	
25) Chloroform	4.268	83	18780	0.439	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
Data File : VV038026.D
Acq On : 12 Nov 2024 18:39
Operator : SY/MD
Sample : P4761-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A0AL0

Quant Time: Nov 13 00:49:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

