

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
 Data File : VV038028.D
 Acq On : 12 Nov 2024 19:28
 Operator : SY/MD
 Sample : P4761-06
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A0AL9

Quant Time: Nov 13 00:49:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	268373	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	276624	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	128798	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	76884	3.291	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.800%
7) Chloroethane-d5	1.526	69	67420	3.541	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.800%
11) 1,1-Dichloroethene-d2	2.053	65	34126	3.334	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.600%
20) 2-Butanone-d5	3.790	46	188878	35.186	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	70.380%
24) Chloroform-d	4.240	84	160537	3.646	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.000%
26) 1,2-Dichloroethane-d4	4.937	65	79279	3.791	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.800%
32) Benzene-d6	4.953	84	304939	3.568	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.400%
36) 1,2-Dichloropropane-d6	5.982	67	98888	3.693	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	73.800%
41) Toluene-d8	7.236	98	258843	3.311	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	66.200%#
43) trans-1,3-Dichloroprop...	7.558	79	30566	2.816	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	56.400%
46) 2-Hexanone-d5	8.018	63	164107	33.822	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	67.640%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	80074	3.615	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	72.400%
66) 1,2-Dichlorobenzene-d4	11.555	152	108668	4.058	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.200%
Target Compounds					Qvalue	
3) Chloromethane	1.214	50	15171	0.559	ug/L	96
13) Acetone	2.114	43	139738	36.519	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
Data File : VV038028.D
Acq On : 12 Nov 2024 19:28
Operator : SY/MD
Sample : P4761-06
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A0AL9

Quant Time: Nov 13 00:49:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

