

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
 Data File : VV038025.D
 Acq On : 12 Nov 2024 18:15
 Operator : SY/MD
 Sample : P4749-20
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GCP93

Quant Time: Nov 13 00:48:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	266635	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	281275	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	130109	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	76722	3.306	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.200%
7) Chloroethane-d5	1.529	69	64294	3.399	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.000%
11) 1,1-Dichloroethene-d2	2.053	65	34412	3.384	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	3.799	46	186711	35.009	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	70.020%
24) Chloroform-d	4.240	84	157767	3.606	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.200%
26) 1,2-Dichloroethane-d4	4.937	65	79633	3.833	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.600%
32) Benzene-d6	4.950	84	310529	3.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.400%
36) 1,2-Dichloropropane-d6	5.982	67	96744	3.553	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	71.000%
41) Toluene-d8	7.236	98	258267	3.249	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.000%#
43) trans-1,3-Dichloroprop...	7.558	79	30934	2.803	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	56.000%
46) 2-Hexanone-d5	8.021	63	166810	33.811	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	67.620%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	83951	3.728	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.600%
66) 1,2-Dichlorobenzene-d4	11.554	152	107002	3.955	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	79.200%#
Target Compounds					Qvalue	
13) Acetone	2.127	43	22980	6.045	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
Data File : VV038025.D
Acq On : 12 Nov 2024 18:15
Operator : SY/MD
Sample : P4749-20
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GCP93

Quant Time: Nov 13 00:48:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

