

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
 Data File : VV038027.D
 Acq On : 12 Nov 2024 19:03
 Operator : SY/MD
 Sample : P4761-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A0AL6

Quant Time: Nov 13 00:49:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	279635	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	289246	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	130979	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	77057	3.166	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.400%
7) Chloroethane-d5	1.529	69	63228	3.187	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	63.800%#
11) 1,1-Dichloroethene-d2	2.053	65	34553	3.240	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.800%
20) 2-Butanone-d5	3.786	46	187501	33.522	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	67.040%
24) Chloroform-d	4.240	84	162571	3.544	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.800%
26) 1,2-Dichloroethane-d4	4.937	65	76319	3.502	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.000%
32) Benzene-d6	4.950	84	304317	3.405	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.200%#
36) 1,2-Dichloropropane-d6	5.982	67	96828	3.458	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	69.200%
41) Toluene-d8	7.236	98	255665	3.128	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	62.600%#
43) trans-1,3-Dichloroprop...	7.558	79	30284	2.668	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	53.400%#
46) 2-Hexanone-d5	8.018	63	153880	30.331	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	60.660%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	81175	3.505	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	70.200%
66) 1,2-Dichlorobenzene-d4	11.554	152	106620	3.915	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	78.400%#
Target Compounds					Qvalue	
13) Acetone	2.114	43	141491	35.488	ug/L	97
25) Chloroform	4.275	83	26043	0.597	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111224\
Data File : VV038027.D
Acq On : 12 Nov 2024 19:03
Operator : SY/MD
Sample : P4761-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A0AL6

Quant Time: Nov 13 00:49:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

