

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
 Data File : VV038044.D
 Acq On : 13 Nov 2024 14:32
 Operator : SY/MD
 Sample : P4761-06
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 13 23:12:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	237191	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	248097	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	114344	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	69291	3.356	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	67.200%	
7) Chloroethane-d5	1.529	69	66989	3.981	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	79.600%	
11) 1,1-Dichloroethene-d2	2.053	65	31059	3.433	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.600%	
20) 2-Butanone-d5	3.793	46	205064	43.223	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.440%	
24) Chloroform-d	4.240	84	170760	4.388	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.800%	
26) 1,2-Dichloroethane-d4	4.937	65	83824	4.535	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.800%	
32) Benzene-d6	4.953	84	302627	3.948	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	79.000%	
36) 1,2-Dichloropropane-d6	5.982	67	99254	4.133	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	82.600%	
41) Toluene-d8	7.239	98	238485	3.402	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	68.000%#	
43) trans-1,3-Dichloroprop...	7.558	79	31681	3.254	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	65.000%	
46) 2-Hexanone-d5	8.021	63	195681	44.967	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.940%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	84454	4.251	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.000%	
66) 1,2-Dichlorobenzene-d4	11.554	152	109270	4.596	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	92.000%	
Target Compounds						
13) Acetone	2.124	43	129130	38.183	ug/L	98
15) Methyl Acetate	2.124	43	127183	17.222	ug/L #	48
35) Methylcyclohexane	5.982	83	25081	0.770	ug/L #	19
37) 1,2-Dichloropropane	5.982	63	10445	0.535	ug/L #	85
61) 1,2,3-Trichloropropane	11.178	75	12672	1.134	ug/L #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
Data File : VV038044.D
Acq On : 13 Nov 2024 14:32
Operator : SY/MD
Sample : P4761-06
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 13 23:12:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

