

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
 Data File : VV038055.D
 Acq On : 13 Nov 2024 19:34
 Operator : SY/MD
 Sample : P4761-15 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 14 01:01:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	273055	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	273326	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	131475	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	71135	2.993	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	59.800%	
7) Chloroethane-d5	1.526	69	68433	3.532	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	70.600%	
11) 1,1-Dichloroethene-d2	2.047	65	34841	3.345	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.000%	
20) 2-Butanone-d5	3.860	46	172186	31.526	ug/L	0.05
Spiked Amount	50.000	Range 40 - 130	Recovery	=	63.060%	
24) Chloroform-d	4.236	84	176491	3.940	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	78.800%	
26) 1,2-Dichloroethane-d4	4.937	65	87416	4.108	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.200%	
32) Benzene-d6	4.950	84	320164	3.791	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	75.800%	
36) 1,2-Dichloropropane-d6	5.982	67	104904	3.965	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	79.400%	
41) Toluene-d8	7.236	98	259708	3.363	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	67.200%#	
43) trans-1,3-Dichloroprop...	7.558	79	36650	3.417	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	68.400%	
46) 2-Hexanone-d5	8.030	63	220533	46.000	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	92.000%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	98662	4.508	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	90.200%	
66) 1,2-Dichlorobenzene-d4	11.554	152	118535	4.336	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	86.800%	
Target Compounds						
35) Methylcyclohexane	5.982	83	27863	0.777	ug/L #	20
42) Toluene	7.333	91	7689	0.085	ug/L	95
48) 2-Hexanone	8.034	43	26295	3.023	ug/L #	68
53) m,p-Xylene	9.079	106	5403	0.145	ug/L	100
61) 1,2,3-Trichloropropane	11.178	75	15528	1.209	ug/L #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
Data File : VV038055.D
Acq On : 13 Nov 2024 19:34
Operator : SY/MD
Sample : P4761-15 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 14 01:01:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

