

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
 Data File : VV038060.D
 Acq On : 13 Nov 2024 21:35
 Operator : SY/MD
 Sample : P4804-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 14 01:02:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	107358	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	112661	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	50484	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	38493	4.119	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	82.400%	
7) Chloroethane-d5	1.529	69	37961	4.984	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.600%	
11) 1,1-Dichloroethene-d2	2.056	65	17912	4.374	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	87.400%	
20) 2-Butanone-d5	3.783	46	139802	65.103	ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	130.200%#	
24) Chloroform-d	4.243	84	88869	5.045	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.000%	
26) 1,2-Dichloroethane-d4	4.940	65	50585	6.047	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	121.000%	
32) Benzene-d6	4.953	84	149416	4.293	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.800%	
36) 1,2-Dichloropropane-d6	5.985	67	51181	4.693	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.800%	
41) Toluene-d8	7.239	98	110848	3.482	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.600%#	
43) trans-1,3-Dichloroprop...	7.564	79	16417	3.714	ug/L	0.01
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.200%	
46) 2-Hexanone-d5	8.017	63	109356	55.339	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	110.680%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	58219	6.454	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	129.000%#	
66) 1,2-Dichlorobenzene-d4	11.554	152	53131	5.062	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.200%	
Target Compounds						
13) Acetone	2.108	43	10893	7.116	ug/L	92
35) Methylcyclohexane	5.985	83	13985	0.946	ug/L #	19
37) 1,2-Dichloropropane	5.982	63	5626	0.635	ug/L #	85
51) Chlorobenzene	8.821	112	5546	0.217	ug/L #	85
61) 1,2,3-Trichloropropane	11.181	75	5394	1.094	ug/L #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
Data File : VV038060.D
Acq On : 13 Nov 2024 21:35
Operator : SY/MD
Sample : P4804-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 14 01:02:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

