

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
 Data File : VV038068.D
 Acq On : 14 Nov 2024 00:47
 Operator : SY/MD
 Sample : P4804-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 14 03:39:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 13 00:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	211582	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	220919	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	96624	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	61353	3.331	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	66.600%	
7) Chloroethane-d5	1.529	69	63336	4.219	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.400%	
11) 1,1-Dichloroethene-d2	2.053	65	26935	3.338	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.800%	
20) 2-Butanone-d5	3.818	46	186751	44.127	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	88.260%	
24) Chloroform-d	4.240	84	157074	4.525	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.400%	
26) 1,2-Dichloroethane-d4	4.937	65	79684	4.833	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.600%	
32) Benzene-d6	4.953	84	259520	3.802	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	76.000%	
36) 1,2-Dichloropropane-d6	5.982	67	94141	4.402	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.000%	
41) Toluene-d8	7.240	98	187553	3.004	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	60.000%#	
43) trans-1,3-Dichloroprop...	7.558	79	27495	3.172	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	63.400%	
46) 2-Hexanone-d5	8.021	63	184207	47.538	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.080%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	85263	4.820	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.400%	
66) 1,2-Dichlorobenzene-d4	11.554	152	84443	4.203	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	84.000%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.278	62	110427	5.140	ug/L	99
8) Chloroethane	1.561	64	1908	0.147	ug/L #	49
14) Carbon disulfide	2.233	76	7530	0.143	ug/L	96
18) trans-1,2-Dichloroethene	2.696	96	4285	0.246	ug/L	95
19) 1,1-Dichloroethane	3.111	63	6611	0.211	ug/L	96
22) cis-1,2-Dichloroethene	3.812	96	27474	1.502	ug/L	95
33) Benzene	5.015	78	17929	0.255	ug/L	100
34) Trichloroethene	5.847	95	2132	0.111	ug/L #	81
35) Methylcyclohexane	5.986	83	24102	0.831	ug/L #	19
61) 1,2,3-Trichloropropane	11.178	75	10483	1.111	ug/L #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111324\
Data File : VV038068.D
Acq On : 14 Nov 2024 00:47
Operator : SY/MD
Sample : P4804-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 14 03:39:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 13 00:39:48 2024
Response via : Initial Calibration

