

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
 Data File : VV038079.D
 Acq On : 14 Nov 2024 09:51
 Operator : SY/MD
 Sample : VSTD0.1
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 15 00:02:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:00:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9864	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9714	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4690	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	967	0.096	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.993	65	557	0.092	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.019	67	747	0.096	ug/L	0.02
8) Toluene-d8	7.269	98	1820	0.091	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	594	0.095	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	953	0.095	ug/L	89
6) Trichloroethene	5.852	95	666	0.094	ug/L	93
9) 1,2-Dibromoethane	8.321	107	464	0.095	ug/L #	93
12) 1,2,3-Trichloropropane	10.214	75	378	0.093	ug/L	93
13) 1,2-Dibromo-3-chloropr...	12.374	75	72	0.091	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
Data File : VV038079.D
Acq On : 14 Nov 2024 09:51
Operator : SY/MD
Sample : VSTD0.1
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 15 00:02:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:00:58 2024
Response via : Initial Calibration





#3

Vinyl chloride

Concen: 0.095 ug/L

RT: 1.280 min Scan# 69

Delta R.T. 0.000 min

Lab File: VV038079.D

Acq: 14 Nov 2024 09:51

Instrument :

MSVOA_V





ClientSampleId :

Tgt Ion: 62 Resp: 953

Ion Ratio Lower Upper

62 100

64 42.7 25.3 47.1



#6

Trichloroethene

Concen: 0.094 ug/L

RT: 5.852 min Scan# 429

Delta R.T. 0.000 min

Lab File: VV038079.D

Acq: 14 Nov 2024 09:51

Tgt Ion: 95 Resp: 666




Ion Ratio Lower Upper


95 100

97 75.1 46.5 86.5

132 95.3 70.3 130.7


130 99.4 73.4 136.4








#9
1,2-Dibromoethane
Concen: 0.095 ug/L
RT: 8.321 min Scan# 51
Delta R.T. 0.016 min
Lab File: VV038079.D
Acq: 14 Nov 2024 09:51

Instrument :
MSVOA_V
ClientSampleId :







Tgt Ion:107 Resp: 464
Ion Ratio Lower Upper
107 100
109 95.7 66.5 123.5
188 35.7 8.2 12.2#



#12
1,2,3-Trichloropropane
Concen: 0.093 ug/L
RT: 10.214 min Scan# 670
Delta R.T. 0.017 min
Lab File: VV038079.D
Acq: 14 Nov 2024 09:51


Tgt Ion: 75 Resp: 378
Ion Ratio Lower Upper
75 100
110 48.1 34.3 51.5
77 30.2 26.6 40.0





#13
1,2-Dibromo-3-chloropropane
Concen: 0.091 ug/L
RT: 12.374 min Scan# 796
Delta R.T. 0.017 min
Lab File: VV038079.D
Acq: 14 Nov 2024 09:51

Instrument :
MSVOA_V
ClientSampleId :



Tgt Ion: 75	Resp: 72
Ion Ratio	Lower Upper
75	100
157	90.0 69.4 104.0
155	84.3 61.3 91.9

