

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
 Data File : VV038083.D
 Acq On : 14 Nov 2024 12:49
 Operator : SY/MD
 Sample : VSTD00224
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002224

Quant Time: Nov 15 00:03:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:00:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	10624	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	10323	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4843	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	23156	2.137	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	14736	2.261	ug/L	-0.02
7) 1,2-Dichloropropane-d6	5.972	67	19002	2.300	ug/L	-0.02
8) Toluene-d8	7.238	98	49279	2.330	ug/L	-0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	14115	2.134	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	21057	1.946	ug/L	98
6) Trichloroethene	5.828	95	14238	1.897	ug/L	99
9) 1,2-Dibromoethane	8.290	107	10113	1.951	ug/L #	98
12) 1,2,3-Trichloropropane	10.197	75	8626	2.045	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	1718	2.099	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
Data File : VV038083.D
Acq On : 14 Nov 2024 12:49
Operator : SY/MD
Sample : VSTD00224
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002224

Quant Time: Nov 15 00:03:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:00:58 2024
Response via : Initial Calibration

