

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
 Data File : VV038085.D
 Acq On : 14 Nov 2024 14:41
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5295

Quant Time: Nov 15 00:23:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9679	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9467	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4607	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	5675	0.566	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 114.000%			
4) 1,2-Dichloroethane-d4	4.969	65	3483	0.574	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 114.000%			
7) 1,2-Dichloropropane-d6	5.995	67	4251	0.554	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 110.000%			
8) Toluene-d8	7.253	98	10843	0.552	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 110.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3330	0.543	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 108.000%			
Target Compounds						
3) Vinyl chloride	1.280	62	5647	0.566	ug/L	Qvalue 100
6) Trichloroethene	5.852	95	3778	0.546	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2543	0.531	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2180	0.540	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.356	75	434	0.544	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
Data File : VV038085.D
Acq On : 14 Nov 2024 14:41
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5295

Quant Time: Nov 15 00:23:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

