

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
 Data File : VV038092.D
 Acq On : 14 Nov 2024 18:08
 Operator : SY/MD
 Sample : P4761-19
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A0AN7

Quant Time: Nov 15 00:33:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9174	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9182	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4649	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	5955	0.626	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	126.000%
4) 1,2-Dichloroethane-d4	4.969	65	3515	0.611	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	122.000%
7) 1,2-Dichloropropane-d6	5.995	67	4199	0.565	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	112.000%
8) Toluene-d8	7.253	98	10708	0.562	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	112.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3314	0.557	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	112.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
Data File : VV038092.D
Acq On : 14 Nov 2024 18:08
Operator : SY/MD
Sample : P4761-19
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A0AN7

Quant Time: Nov 15 00:33:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

