

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
 Data File : VV038093.D
 Acq On : 14 Nov 2024 18:32
 Operator : SY/MD
 Sample : P4803-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A0AM1

Quant Time: Nov 15 00:33:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9190	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9107	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4618	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6002	0.630	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 126.000%	
4) 1,2-Dichloroethane-d4	4.969	65	3543	0.615	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 122.000%	
7) 1,2-Dichloropropane-d6	5.995	67	4213	0.571	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 114.000%	
8) Toluene-d8	7.253	98	10724	0.568	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 114.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3409	0.578	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 116.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111424\
Data File : VV038093.D
Acq On : 14 Nov 2024 18:32
Operator : SY/MD
Sample : P4803-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A0AM1

Quant Time: Nov 15 00:33:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

