

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111524\
 Data File : VV038127.D
 Acq On : 15 Nov 2024 10:00
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5298

Quant Time: Nov 15 22:10:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9934	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9623	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4667	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	5617	0.546	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 110.000%			
4) 1,2-Dichloroethane-d4	4.970	65	3396	0.545	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 110.000%			
7) 1,2-Dichloropropane-d6	5.996	67	4094	0.525	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 106.000%			
8) Toluene-d8	7.254	98	10594	0.531	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 106.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3269	0.525	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 104.000%			
Target Compounds						
3) Vinyl chloride	1.273	62	5564	0.544	ug/L	97
6) Trichloroethene	5.852	95	3685	0.523	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2454	0.504	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2076	0.508	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.357	75	410	0.507	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111524\
Data File : VV038127.D
Acq On : 15 Nov 2024 10:00
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5298

Quant Time: Nov 15 22:10:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

