

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV111524\  
 Data File : VV038136.D  
 Acq On : 15 Nov 2024 14:59  
 Operator : SY/MD  
 Sample : P4761-14  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 A0AK5

Quant Time: Nov 15 22:11:18 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM111424.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Fri Nov 15 00:15:53 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.566  | 114   | 11241    | 0.500    | ug/L  | 0.00     |
| 5) Chlorobenzene-d5           | 8.800  | 117   | 11099    | 0.500    | ug/L  | 0.00     |
| 11) 1,4-Dichlorobenzene-d4    | 11.191 | 152   | 5467     | 0.500    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 2) Vinyl Chloride-d3          | 1.273  | 65    | 6294     | 0.540    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 40 - 130 | Recovery | =     | 108.000% |
| 4) 1,2-Dichloroethane-d4      | 4.969  | 65    | 3372     | 0.478    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =     | 96.000%  |
| 7) 1,2-Dichloropropane-d6     | 5.995  | 67    | 4060     | 0.452    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 60 - 140 | Recovery | =     | 90.000%  |
| 8) Toluene-d8                 | 7.253  | 98    | 11032    | 0.479    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =     | 96.000%  |
| 10) 1,1,2,2-Tetrachloroeth... | 10.146 | 84    | 3336     | 0.464    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 65 - 120 | Recovery | =     | 92.000%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV111524\  
Data File : VV038136.D  
Acq On : 15 Nov 2024 14:59  
Operator : SY/MD  
Sample : P4761-14  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
A0AK5

Quant Time: Nov 15 22:11:18 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM111424.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Fri Nov 15 00:15:53 2024  
Response via : Initial Calibration

