

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111623\
 Data File : VV032963.D
 Acq On : 16 Nov 2023 09:28
 Operator : SY/MD
 Sample : VV1116MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK218

Quant Time: Nov 17 00:56:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 17 00:55:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	435514	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	440733	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	216633	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	156975	42.616	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.240%
7) Chloroethane-d5	1.532	69	137386	45.017	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.040%
11) 1,1-Dichloroethene-d2	2.059	65	70660	44.029	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.060%
21) 2-Butanone-d5	3.780	46	129945	82.302	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.300%
24) Chloroform-d	4.246	84	298652	46.933	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.860%
26) 1,2-Dichloroethane-d4	4.937	65	181814	46.957	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.920%
32) Benzene-d6	4.957	84	572000	47.585	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.180%
36) 1,2-Dichloropropane-d6	5.985	67	170503	48.798	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.600%
41) Toluene-d8	7.243	98	510084	46.026	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.060%
43) trans-1,3-Dichloroprop...	7.551	79	85354	46.015	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.040%
47) 2-Hexanone-d5	8.021	63	109396	85.462	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.460%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	227783	44.503	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	215763	51.257	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111623\
Data File : VV032963.D
Acq On : 16 Nov 2023 09:28
Operator : SY/MD
Sample : VV1116MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK218

Quant Time: Nov 17 00:56:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 17 00:55:52 2023
Response via : Initial Calibration

