

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111624\
 Data File : VV038153.D
 Acq On : 16 Nov 2024 10:17
 Operator : SY/MD
 Sample : VV1116WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 17 23:35:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9455	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9376	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4550	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	5580	0.569	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 114.000%			
4) 1,2-Dichloroethane-d4	4.970	65	3486	0.588	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 118.000%			
7) 1,2-Dichloropropane-d6	5.996	67	4102	0.540	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 108.000%			
8) Toluene-d8	7.254	98	10423	0.536	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3333	0.549	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 110.000%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111624\
Data File : VV038153.D
Acq On : 16 Nov 2024 10:17
Operator : SY/MD
Sample : VV1116WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 17 23:35:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

