

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111624\
 Data File : VV038152.D
 Acq On : 16 Nov 2024 09:05
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5300

Quant Time: Nov 17 23:34:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	10398	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.784	117	10334	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.191	152	4877	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6075	0.564	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	112.000%
4) 1,2-Dichloroethane-d4	4.945	65	3445	0.528	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	5.995	67	4485	0.536	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.000%
8) Toluene-d8	7.238	98	11291	0.527	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
10) 1,1,2,2-Tetrachloroeth...	10.145	84	3349	0.500	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	100.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	5769	0.539	ug/L	# 39
6) Trichloroethene	5.852	95	3801	0.503	ug/L	98
9) 1,2-Dibromoethane	8.305	107	2444	0.468	ug/L	98
12) 1,2,3-Trichloropropane	10.197	75	2060	0.482	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.356	75	416	0.493	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111624\
Data File : VV038152.D
Acq On : 16 Nov 2024 09:05
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5300

Quant Time: Nov 17 23:34:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

