

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111624\
 Data File : VV038159.D
 Acq On : 16 Nov 2024 13:26
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5301

Quant Time: Nov 17 23:36:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9425	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9326	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4578	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	5260	0.538	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 108.000%			
4) 1,2-Dichloroethane-d4	4.969	65	3313	0.560	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 112.000%			
7) 1,2-Dichloropropane-d6	5.995	67	3867	0.512	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 102.000%			
8) Toluene-d8	7.253	98	9647	0.499	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 100.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3337	0.553	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 110.000%			
Target Compounds						
3) Vinyl chloride	1.280	62	5180	0.533	ug/L #	32
6) Trichloroethene	5.852	95	3297	0.483	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2341	0.496	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2051	0.512	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.356	75	405	0.511	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111624\
Data File : VV038159.D
Acq On : 16 Nov 2024 13:26
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5301

Quant Time: Nov 17 23:36:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

