

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
 Data File : VV032981.D
 Acq On : 17 Nov 2023 02:58
 Operator : SY/MD
 Sample : 05341-15
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0QH8

Quant Time: Nov 17 04:58:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 00:51:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	174181	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	179496	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	75539	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	82502	5.565	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.200%
7) Chloroethane-d5	1.532	69	69970	4.923	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.400%
11) 1,1-Dichloroethene-d2	2.059	65	42944	6.570	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	131.400%#
20) 2-Butanone-d5	3.793	46	87830	44.466	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.940%
24) Chloroform-d	4.252	84	130618	5.124	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
26) 1,2-Dichloroethane-d4	4.944	65	58070	5.160	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	4.963	84	247414	5.243	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.800%
36) 1,2-Dichloropropane-d6	5.992	67	68723	5.130	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.600%
41) Toluene-d8	7.246	98	210084	4.732	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
43) trans-1,3-Dichloroprop...	7.561	79	23983	4.593	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.800%
46) 2-Hexanone-d5	8.027	63	79308	44.421	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.840%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	45894	4.644	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	75127	5.715	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
Data File : VV032981.D
Acq On : 17 Nov 2023 02:58
Operator : SY/MD
Sample : 05341-15
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0QH8

Quant Time: Nov 17 04:58:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 00:51:54 2023
Response via : Initial Calibration

