

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
 Data File : VV032987.D
 Acq On : 17 Nov 2023 05:23
 Operator : SY/MD
 Sample : 05341-21
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0QM2

Quant Time: Nov 17 06:31:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 00:51:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	187201	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	197302	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	83093	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	96662	6.066	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 121.400%		
7) Chloroethane-d5	1.532	69	84216	5.513	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 110.200%		
11) 1,1-Dichloroethene-d2	2.060	65	41708	5.937	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 118.800%		
20) 2-Butanone-d5	3.793	46	108116	50.929	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 101.860%		
24) Chloroform-d	4.249	84	142686	5.208	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 104.200%		
26) 1,2-Dichloroethane-d4	4.944	65	64250	5.312	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 106.200%		
32) Benzene-d6	4.963	84	276520	5.331	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 106.600%		
36) 1,2-Dichloropropane-d6	5.992	67	81920	5.563	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 111.200%		
41) Toluene-d8	7.246	98	232127	4.757	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 95.200%		
43) trans-1,3-Dichloroprop...	7.558	79	27632	4.814	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 96.200%		
46) 2-Hexanone-d5	8.024	63	89526	45.619	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 91.240%		
56) 1,1,2,2-Tetrachloroeth...	10.156	84	47662	4.388	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 87.800%		
66) 1,2-Dichlorobenzene-d4	11.561	152	83172	5.752	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 115.000%		
Target Compounds						Qvalue
13) Acetone	2.121	43	5936	3.011	ug/L	88
42) Toluene	7.326	91	10535	0.166	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
Data File : VV032987.D
Acq On : 17 Nov 2023 05:23
Operator : SY/MD
Sample : 05341-21
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0QM2

Quant Time: Nov 17 06:31:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 00:51:54 2023
Response via : Initial Calibration

