

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111723\
 Data File : VW032989.D
 Acq On : 17 Nov 2023 06:11
 Operator : SY/MD
 Sample : 05389-17
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCGE6

Quant Time: Nov 17 06:31:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 00:51:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	192977	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	203561	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	86207	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	91543	5.573	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.400%
7) Chloroethane-d5	1.532	69	82205	5.220	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.400%
11) 1,1-Dichloroethene-d2	2.060	65	41371	5.713	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	114.200%
20) 2-Butanone-d5	3.799	46	114555	52.347	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.700%
24) Chloroform-d	4.249	84	147713	5.230	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%
26) 1,2-Dichloroethane-d4	4.944	65	66835	5.361	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.200%
32) Benzene-d6	4.963	84	287113	5.365	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.400%
36) 1,2-Dichloropropane-d6	5.989	67	87197	5.740	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.800%
41) Toluene-d8	7.246	98	241197	4.791	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
43) trans-1,3-Dichloroprop...	7.558	79	28582	4.826	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.600%
46) 2-Hexanone-d5	8.027	63	95618	47.225	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.440%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	52641	4.697	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	11.564	152	80096	5.339	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%
Target Compounds						
					Qvalue	
14) Carbon disulfide	2.243	76	39796	0.846	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
Data File : VV032989.D
Acq On : 17 Nov 2023 06:11
Operator : SY/MD
Sample : 05389-17
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCGE6

Quant Time: Nov 17 06:31:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 00:51:54 2023
Response via : Initial Calibration

