

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111723\
 Data File : VW032992.D
 Acq On : 17 Nov 2023 07:23
 Operator : SY/MD
 Sample : 05333-17
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCG41

Quant Time: Nov 17 22:37:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 00:51:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	130469	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	138369	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	58910	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	64887	5.843	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	116.800%	
7) Chloroethane-d5	1.532	69	48910	4.594	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.800%	
11) 1,1-Dichloroethene-d2	2.060	65	23634	4.827	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	96.600%	
20) 2-Butanone-d5	3.786	46	132138	89.311	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	178.620%#	
24) Chloroform-d	4.253	84	105654	5.534	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	110.600%	
26) 1,2-Dichloroethane-d4	4.947	65	54423	6.456	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	129.200%	
32) Benzene-d6	4.963	84	197468	5.428	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	108.600%	
36) 1,2-Dichloropropane-d6	5.992	67	65253	6.319	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	126.400%	
41) Toluene-d8	7.246	98	154717	4.521	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.400%	
43) trans-1,3-Dichloroprop...	7.558	79	23107	5.740	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	114.800%	
46) 2-Hexanone-d5	8.024	63	100937	73.339	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	146.680%#	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	52639	6.910	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	138.200%#	
66) 1,2-Dichlorobenzene-d4	11.564	152	56702	5.531	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	110.600%	
Target Compounds						
13) Acetone	2.111	43	14462	10.524	ug/L	98
16) Methylene chloride	2.449	84	11142	0.854	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
Data File : VV032992.D
Acq On : 17 Nov 2023 07:23
Operator : SY/MD
Sample : 05333-17
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCG41

Quant Time: Nov 17 22:37:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 00:51:54 2023
Response via : Initial Calibration

