

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111723\
 Data File : VW032994.D
 Acq On : 17 Nov 2023 08:11
 Operator : SY/MD
 Sample : 05333-19
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCG52

Quant Time: Nov 17 22:38:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 00:51:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	209196	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	217893	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	91035	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	102709	5.768	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.400%
7) Chloroethane-d5	1.532	69	77509	4.540	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.800%
11) 1,1-Dichloroethene-d2	2.059	65	38716	4.931	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.600%
20) 2-Butanone-d5	3.793	46	124326	52.408	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.820%
24) Chloroform-d	4.252	84	160879	5.255	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.200%
26) 1,2-Dichloroethane-d4	4.944	65	68973	5.103	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.000%
32) Benzene-d6	4.963	84	313554	5.474	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.400%
36) 1,2-Dichloropropane-d6	5.989	67	95556	5.876	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.600%
41) Toluene-d8	7.246	98	248998	4.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
43) trans-1,3-Dichloroprop...	7.558	79	29370	4.633	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.600%
46) 2-Hexanone-d5	8.027	63	104883	48.393	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.780%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	59304	4.944	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	86333	5.450	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
Data File : VV032994.D
Acq On : 17 Nov 2023 08:11
Operator : SY/MD
Sample : 05333-19
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCG52

Quant Time: Nov 17 22:38:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 00:51:54 2023
Response via : Initial Calibration

