

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111723\
 Data File : VW032991.D
 Acq On : 17 Nov 2023 06:59
 Operator : SY/MD
 Sample : 05389-10
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCG58

Quant Time: Nov 17 07:21:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 00:51:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	225696	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	229774	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	96138	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	105277	5.480	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.600%
7) Chloroethane-d5	1.532	69	78231	4.248	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.000%
11) 1,1-Dichloroethene-d2	2.060	65	37634	4.443	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.800%
20) 2-Butanone-d5	3.796	46	124202	48.528	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.060%
24) Chloroform-d	4.252	84	157492	4.768	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	4.947	65	67350	4.619	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
32) Benzene-d6	4.963	84	314655	5.209	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%
36) 1,2-Dichloropropane-d6	5.992	67	95647	5.578	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.600%
41) Toluene-d8	7.246	98	258012	4.540	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.558	79	27981	4.186	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.027	63	104407	45.683	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.360%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	56654	4.479	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	85977	5.139	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111723\
Data File : VV032991.D
Acq On : 17 Nov 2023 06:59
Operator : SY/MD
Sample : 05389-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCG58

Quant Time: Nov 17 07:21:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 00:51:54 2023
Response via : Initial Calibration

