

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111821\
 Data File : VW023602.D
 Acq On : 18 Nov 2021 14:24
 Operator : SY/MD
 Sample : M4627-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4669

Quant Time: Nov 19 02:13:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 02:11:08 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	116691	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	115984	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55911	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	38054	5.206	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 104.200%		
7) Chloroethane-d5	1.568	69	30513	5.121	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 102.400%		
11) 1,1-Dichloroethene-d2	2.108	63	52262	3.819	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 76.400%		
20) 2-Butanone-d5	3.892	46	50398	40.017	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 80.040%		
24) Chloroform-d	4.349	84	69861	4.484	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 89.600%		
26) 1,2-Dichloroethane-d4	5.037	65	34080	4.865	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 97.200%		
32) Benzene-d6	5.050	84	133546	4.488	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 89.800%		
36) 1,2-Dichloropropane-d6	6.069	67	39726	4.535	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 90.600%		
41) Toluene-d8	7.317	98	116327	4.171	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 83.400%		
43) trans-1,3-Dichloroprop...	7.625	79	14274	4.297	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 86.000%		
46) 2-Hexanone-d5	8.091	63	53857	44.067	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 88.140%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	27850	4.421	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 88.400%		
66) 1,2-Dichlorobenzene-d4	11.625	152	46153	4.957	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 99.200%		
Target Compounds						Qvalue
5) Vinyl chloride	1.310	62	55320	5.726	ug/L	97
13) Acetone	2.178	43	10053	13.064	ug/L	85
14) Carbon disulfide	2.294	76	8400	0.320	ug/L	99
16) Methylene chloride	2.510	84	4456	0.439	ug/L	97
18) trans-1,2-Dichloroethene	2.764	96	4993	0.584	ug/L	95
19) 1,1-Dichloroethane	3.191	63	8889	0.615	ug/L	95
22) cis-1,2-Dichloroethene	3.915	96	31062	3.773	ug/L #	87
33) Benzene	5.108	78	6166	0.190	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
Data File : VV023602.D
Acq On : 18 Nov 2021 14:24
Operator : SY/MD
Sample : M4627-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4669

Quant Time: Nov 19 02:13:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 02:11:08 2021
Response via : Initial Calibration

