

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
 Data File : VV023609.D
 Acq On : 18 Nov 2021 17:16
 Operator : SY/MD
 Sample : VV1118WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK259

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/19/2021
 Supervised By : Mahesh Dadoda 11/19/2021

Quant Time: Nov 19 03:52:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 03:51:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	121151	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	116966	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54274	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	37389	4.926	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	98.600%	
7) Chloroethane-d5	1.568	69	29399	4.753	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.000%	
11) 1,1-Dichloroethene-d2	2.108	63	51317	3.612	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	72.200%	
20) 2-Butanone-d5	3.899	46	56558m	43.255	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.500%	
24) Chloroform-d	4.346	84	67022	4.144	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.800%	
26) 1,2-Dichloroethane-d4	5.031	65	32166	4.422	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.400%	
32) Benzene-d6	5.047	84	130352	4.343	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
36) 1,2-Dichloropropane-d6	6.066	67	37588	4.255	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	85.000%	
41) Toluene-d8	7.317	98	115020	4.090	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.800%	
43) trans-1,3-Dichloroprop...	7.625	79	13149	3.925	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.600%	
46) 2-Hexanone-d5	8.092	63	50072	40.626	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	81.260%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26615	4.189	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	83.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	44154	4.886	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
Data File : VV023609.D
Acq On : 18 Nov 2021 17:16
Operator : SY/MD
Sample : VV1118WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK259

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/19/2021
Supervised By : Mahesh Dadoda 11/19/2021

Quant Time: Nov 19 03:52:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 03:51:44 2021
Response via : Initial Calibration

